

Certificate of Analysis

Product Name:	Folic Acid (Folate) Assay Diluent
Catalog Number:	FA501
Lot Number:	20210406F
Form:	Liquid
Buffer:	Tris buffer
pH:	7.7~8.3
Preservative:	0.05% ProClin-300
Appearance:	Clear
Application:	Separate folate from the folate-binding protein in serum/plasma. Recommended adding ratio (Sample: Diluent = 1:1). The amount of sample added after mixing depends on your situation, generally not more than 50 μ l.
Storage:	Store at room temperature for two years.
Note:	For diagnostic use only. Not intended for therapeutic use. Standard laboratory practices should be followed when handling.
Quality Department:	Date: 2021.04.07