

Certificate of Analysis

Product Name:	Folic Acid (Folate) Assay Diluent, 2-components
Catalog Number:	FA502 (FA502-1 component R1; FA502-2 component R2)
Lot Number:	20210527A
Form:	Liquid
Buffer:	0.01M PBS
pH	FA502-1: 4.8-5.2 FA502-2: 12.5-13.0
Preservative:	0.05% ProClin-300
Appearance:	Clear
Application:	Separate folate from the folate-binding protein in serum/plasma. Recommended adding ratio (R1:R2 = 5:1 ~ 10:1, Sample : (R1+R2) \approx 1 : 1). React at 37°C or room temperature for 8-15 minutes.
Storage:	Store at room temperature for two years.
Note:	For diagnostic use only. Not intended for therapeutic use. Standard laboratory practices should be followed when handling.

Quality Department:

Date: 2021.05.28